

Note: This is a reference cited in *AP 42, Compilation of Air Pollutant Emission Factors, Volume I Stationary Point and Area Sources*. AP42 is located on the EPA web site at www.epa.gov/ttn/chief/ap42/

The file name refers to the reference number, the AP42 chapter and section. The file name "ref02_c01s02.pdf" would mean the reference is from AP42 chapter 1 section 2. The reference may be from a previous version of the section and no longer cited. The primary source should always be checked.

Background Report Reference

AP-42 Section Number: 10.6.1

Background Report Section: 4

Reference Number: 26

Title: Letter and Attachment from M. Wood and L. Kessari, USEPA, Washington, DC Calibration Drift Results for September 16, 1994 test on CO and Opacity Monitoring Systems on RTO Stack Pacific-Corp., OSB Plant, Chilco, Idaho, October 28, 1994



Louisiana-Pacific Corporation

4-26)

Route 8, Box 8263
Hayward, Wisconsin 54843
715/634-3332
FAX: 715/634-5117

October 28, 1994

Mr. Michael F. Wood, Director
Mr. Laxmi Kesari
Multi Media Enforcement & Strategic
Planning Division U.S. EPA
Aerials Rios Boulevard
1200 Pennsylvania N. W., 7th Floor
Room 7120A
Washington D.C. 20004

RE: Clean Air Enforcement Action - United States v. Louisiana Pacific Corporation
No. CV 93-0869 (W.D. La.)

Subject: Continuous Monitoring Equipment Certification: L-P Chilco OSB
CO and Opacity Drift Test Results.

Gentlemen:

Enclosed please find the calibration drift (CO) test results for the carbon monoxide (CO) and opacity continuous monitoring systems installed at the RTO exhaust stack at Louisiana Pacific's OSB plant in Chilco, Idaho. These are submitted in conjunction with the September 30, 1994 (request for an extension) letter from Elizabeth Smith, Director of Environmental Affairs - Louisiana Pacific Corporation, Portland, Oregon. Attach these test results to your copy of the AmTest Air Quality, Inc. Performance Specification Test (PST) Report dated September 29, 1994.

Mr. Michael Wood/Laxmi Kesari
October 28, 1994
Page 2

If you have any questions or need additional information please contact me.

Sincerely,



Robert W. Schultz
Environmental Engineer
Hayward, WI Engineering Office

cc w/encls: Mike Razzeto - Chilco Idaho
Mike Simon/Ron Hill, Division of Environmental Quality, Boise, ID
Liz Smith, Director of Environmental Affairs, Portland, OR


w/out encls: Julie Domike U.S. EPA, Office of Civil Enforcement, Washington, D.C.
U.S. EPA, Region 10, Seattle, WA
Jim Evensen - Hayden Lake
Keith Seelig - Hayward
Bert Krages - Portland
Norm Radford Jr. - Vinson & Elkins, Dallas, TX



AmTest-Air Quality, Inc.
30545 S.E. 84th St., #5
Preston, WA 98050
Office: (206) 222-7746
FAX: (206) 222-7849

October 26, 1994

Mr. Bob Schultz
Louisiana-Pacific Corporation
Route 8, Box 8263
Hayward, WI 54843



Dear Bob:

Please find enclosed the calibration drift (CD) test results which are part of the performance specification test (PST) performed for the carbon monoxide (CO) continuous emission monitoring system (CEMS) installed at the RTO exhaust stack at Louisiana Pacific's facility in Chilco, Idaho.

Performance Specification 4, Section 2.2, states that the calibration drift of the CO analyzer must not drift or deviate from the reference value by more than 5% of the span value for 6 out of 7 days. The maximum zero drift for the CO analyzer was 0.56% of span. The maximum calibration drift for the CO analyzer was 2.0% of span.

Four (4) copies of a complete data packet with a computer printout of results, example calculations and back-up CEMS data is enclosed. Copies of this information should be forwarded to the appropriate parties.

Please don't hesitate to call our office if you have any questions or require additional information.

Sincerely,
Am Test-Air Quality, Inc.

A handwritten signature in cursive script that reads "Angela F. Blaisdell".

Angela F. Blaisdell
Sr. Technical Writer

AFB/amb
Enclosure



24-HOUR ZERO AND CALIBRATION DRIFT TEST
AM TEST - AIR QUALITY, INC.

FILE NAME: R316\LPCOCO
CLIENT: Louisiana-Pacific
LOCATION: Chilco, Idaho
MONITOR LOCATION: RTO Exhaust Stack
MONITOR MANUFACTURER: Thermol Environmental Model 48
TYPE OF MONITOR: Carbon Monoxide Analyzer
INSTRUMENT SPAN: 50 ppm CO

ZERO (LOW LEVEL) DRIFT (ppm CARBON MONOXIDE)

DATE	TIME	CAL. TANK VALUE ppm	MONITOR READING ppm	ABSOLUTE DIFFERENCE ppm	PERCENT OF SPAN
8/23/94	06:09		0.09		
8/24/94	06:09	0.00	0.14	0.14	0.28
8/25/94	06:09	0.00	0.28	0.28	0.56
8/26/94	06:09	0.00	0.05	0.05	0.10
8/27/94	06:09	0.00	-0.13	0.13	0.26
8/28/94	06:09	0.00	0.10	0.10	0.20
8/29/94	06:09	0.00	-0.10	0.10	0.20
8/30/94	06:09	0.00	-0.06	0.06	0.12

SPAN (HIGH LEVEL) DRIFT (ppm CARBON MONOXIDE)

DATE	TIME	CAL. TANK VALUE ppm	MONITOR READING ppm	ABSOLUTE DIFFERENCE ppm	PERCENT OF SPAN
8/23/94	06:09		46.15		
8/24/94	06:09	46.00	46.23	0.23	0.46
8/25/94	06:09	46.00	47.00	1.00	2.00
8/26/94	06:09	46.00	46.05	0.05	0.10
8/27/94	06:09	46.00	45.87	0.13	0.26
8/28/94	06:09	46.00	46.09	0.09	0.18
8/29/94	06:09	46.00	45.85	0.15	0.30
8/30/94	06:09	46.00	45.86	0.14	0.28

CRITERIA: Calibration should not drift from the calibration gas value by more than 5 percent of the span value for 6 out of 7 days (2.5 ppm).

**24-HOUR ZERO AND CALIBRATION DRIFT
CONTINUOUS EMISSION MONITORING SYSTEM PERFORMANCE TEST**

Client Louisiana-Pacific Location Chillico, ID
 Monitor Manufacturer TECO Model No. 48
 Type of Monitor (SO₂, NO_x, HC, etc.) CO
 Unit of Measurement (% , ppm, wet or dry basis) ppm
 Measurement Principle (NDIR, chemiluminescent, etc) NDIR
 Span Value of the Instrument 50

ZERO DRIFT TEST						
Day	Date	Time	Calibration Value	Monitor Value	Difference	Percent of Span Value
0	8/23	06:09	0.00	0.09	N/A	N/A
1	8/24			0.14	0.14	0.28
2	8/25			0.28	0.28	0.56
3	8/26			0.05	0.05	0.10
4	8/27			-0.13	0.13	0.26
5	8/28			0.10	0.10	0.20
6	8/29			-0.10	0.10	0.20
7	8/30	↓	↓	-0.11	0.11	0.12

CALIBRATION DRIFT TEST						
Day	Date	Time	Calibration Value	Monitor Value	Difference	Percent of Span Value
0	8/23	06:09	46.00	46.15	N/A	N/A
1	8/24			46.23	0.23	0.46
2	8/25			47.00	1.00	2.00
3	8/26			46.05	0.05	0.10
4	8/27			45.87	0.13	0.26
5	8/28			46.09	0.09	0.18
6	8/29			45.85	0.15	0.30
7	8/30	↓	↓	45.86	0.14	0.28



EXAMPLE CALCULATION OF ZERO AND CALIBRATION DRIFT

Client: Louisiana - Pacific

Location: Chilco, Idaho

Site Location: RTO Exhaust Stack

Analyzer Type: carbon monoxide

Percent of Span Value - Zero Drift

Day: 4 Date: 8/26/94

| Monitor Value - Calibration Value | = | Absolute Difference |

| 0.05 - 0.0 | = | 0.05 |

$\frac{| \text{Absolute Difference} |}{\text{Span Value}} * 100\% = \text{Percent of Span Value}$

| $\frac{0.05}{50}$ | * 100% = 0.1 %

Percent of Span Value - Calibration Drift

Day: 4 Date: 8/26/94

| Monitor Value - Calibration Value | = | Absolute Difference |

| 46.05 - 46.0 | = | 0.05 |

$\frac{| \text{Absolute Difference} |}{\text{Span Value}} * 100\% = \text{Percent of Span Value}$

| $\frac{0.05}{50}$ | * 100% = 0.1 %

SENT BY:LOUISIANA-PACIFIC

: 9-28-94 : 5:06PM :

ENGINEERING DEPT.→

206 222 7849:#10

RTO Stack NORMAL CALIBRATION REPORT

LOUISIANA PACIFIC

06:09:01

08-23-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.29	0.09	0.19	0.59
CO Span	46.00	48.35	46.15	0.30	96.70

RTO Stack NORMAL CALIBRATION REPORT

LOUISIANA PACIFIC

06:09:01

08-24-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.44	0.14	0.28	0.88
CO Span	46.00	48.60	46.23	0.47	97.19

SENT BY:LOUISIANA-PACIFIC

; 9-28-94 ; 5:05PM ;

ENGINEERING DEPT. →

206 222 7849:# 8

RTO Stack NORMAL CALIBRATION REPORT

LOUISIANA PACIFIC

06:09:01

08-25-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.73	0.28	0.56	1.46
CO Span	46.00	49.65	47.00	2.01	99.29

=====

RTO Stack NORMAL CALIBRATION REPORT

LOUISIANA PACIFIC

06:09:01

08-26-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.78	0.05	0.09	1.56
CO Span	46.00	49.69	46.05	0.09	99.39

RTO Stack NORMAL CALIBRATION REPORT
LOUISIANA PACIFIC
06109101 08-27-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.65	-0.13	-0.25	1.29
CO Span	46.00	49.56	45.87	-0.25	99.12

RTO Stack NORMAL CALIBRATION REPORT

LOUISIANA PACIFIC

06:09:00

08-28-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.76	0.10	0.21	1.51
CO Span	46.00	49.66	46.09	0.18	99.32

=====

RTO Stack NORMAL CALIBRATION REPORT
LOUISIANA PACIFIC
06109:01 08-29-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.65	-0.10	-0.21	1.29
CO Span	46.00	49.50	45.85	-0.30	99.00

=====

RTO Stack NORMAL CALIBRATION REPORT
LOUISIANA PACIFIC

06:09:00

08-30-1994

SAMPLE	CAL GAS CONCENTRATION	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT FULL SCALE
CO Zero	0.00	0.59	-0.06	-0.11	1.17
CO Span	46.00	49.35	45.86	-0.28	98.71

10-25-94

USI MODEL 500C ZERO & CALIBRATION DRIFT

PERSON CONDUCTING TEST BOB SCHULTZ
 AFFILIATION LOUISIANA-PACIFIC CORP./ENG.
 ANALYZER MANUFACTURER 0993724 UNITED SCIENCES, INC.
 SERIAL NUMBER
 TEST DATES 10-8-94 THROUGH 10-15-94
 LOCATION LOUISIANA-PACIFIC CORP. - CHILCO, ID - RTO STACK
 MONITOR PATHLENGTH - L1
 EMISSION OUTLET PATHLENGTH - L2
 MONITORING SYSTEM OUTPUT PATHLENGTH CORRECTED: YES NO XX
 UPSCALE CALIBRATION VALUE: ACTUAL OPTICAL DENSITY (OPACITY) - (45.94%)
 PATH ADJUSTED OPTICAL DENSITY (OPACITY) - (45.94%)

PERCENT OPACITY																		
DATE	TIME		ZERO READING		ZERO DRIFT C = B - A	Z E R O	A D J U S T E D	UPSCALE CALIBRATION READING		3, UPSCALE DRIFT F = E - D	2.CALI- BRATION DRIFT G = F - C	S P A C E	LENS CLEANED	ALIGNMENT		ACCUM. DIRT COMP IN % OPACITY		
	BEGIN	END	INITIAL A	FINAL B				INITIAL D	FINAL E								CHECK	
T0 10-8	0834	0900	--	(-2.15) -3.70	--	NO	--	--	45.94	--	--	NO	NO	NO	NO	-3.70		
T1 10-9	0834	0900	(-2.15) -3.70	(-2.03) -3.59	+0.11	NO	45.71	45.94	-0.24	-0.24	-0.24	NO	NO	NO	NO	-3.59		
T2 10-10	0834	0900	(-2.03) -3.59	(-1.78) -3.35	+0.24	NO	45.87	45.94	-0.07	-0.07	-0.07	NO	NO	NO	NO	-3.35		
T3 10-11	0834	0900	(-1.78) -3.35	(-1.21) -3.38	-0.03	NO	45.79	45.94	-0.16	-0.16	-0.16	NO	NO	NO	NO	-3.38		
T4 10-12	0834	0900	(-1.21) -3.38	(-1.90) -3.46	-0.08	NO	45.76	45.94	-0.18	-0.18	-0.18	NO	NO	NO	NO	-3.46		
T5 10-13	0834	0900	(-1.90) -3.46	(-1.83) -3.39	+0.07	NO	45.62	45.94	-0.32	-0.32	-0.32	NO	NO	NO	NO	-3.39		
T6 10-14	0834	0900	(-1.83) -3.39	(-1.81) -3.38	+0.01	NO	45.83	45.94	-0.11	-0.11	-0.11	NO	NO	NO	NO	-3.38		
T7 10-15	0834	0900	(-1.81) -3.38	(-1.90) -3.46	-0.08	NO	45.92	45.94	-0.02	-0.02	-0.02	NO	NO	NO	NO	-3.46		
ARITHMETIC MEAN					+0.034					-0.157								
CONFIDENCE INTERVAL					0.066					0.373								
ZERO DRIFT					0.100					0.530								

1. Accumulated dirt compensation recorded as final zero reading. Output device indication in parenthesis.
2. Zero was adjusted automatically prior to upscale check, therefore C = 0.
3. Span drift = ADJ RDG - CAL STD per Graseby-STI.

10-25-74
R. Schulte
18 of 4

CONTINUOUS EMISSION MONITORING SYSTEM (CEMS)
RELATIVE ACCURACY TEST AUDIT (RATA)

LOUISIANA-PACIFIC CORP. Location CHILCO, ID RTO STACK
 Monitor Manufacturer UNITED SCIENCES, INC. Model No. 500C
 Type of Monitor (SO₂, NO_x, HC, etc.) OPACITY
 Unit of Measurement (% , ppm, wet or dry basis) %
 Emission Standard (provide units) 20%
 Measurement Principle (NDIR, chemiluminescent, etc.) _____
 Monitor System Type (extractive, in-situ) _____
 Range of Instrument _____ Span Gas Concentration _____

RELATIVE ACCURACY TEST <u>ZERO DRIFT</u>						
Test No.	Date	Time	INITIAL Reference Method Value	FINAL CEMS Value	Difference X _i	Difference X _i ²
1	10-08	0834-0900	-	-3.70	--	--
2	10-09		-3.70	-3.59	+0.11	0.0121
3	10-10		-3.59	-3.35	+0.24	0.0576
4	10-11		-3.35	-3.38	-0.03	0.0009
5	10-12		-3.38	-3.46	-0.08	0.0064
6	10-13		-3.46	-3.39	+0.07	0.0049
7	10-14		-3.39	-3.38	+0.01	0.0001
8	10-15	X	-3.38	-3.46	-0.08	0.0064
9						
10						
11						
12						
Mean			a		b ± 0.034	
Sum of the Differences						c 0.0884
(Sum of the Differences) ²					d 0.0576	

2 of 4

EXAMPLE CALCULATION OF
CONFIDENCE INTERVAL AND RELATIVE ACCURACY
CONTINUOUS EMISSION MONITORING SYSTEM PERFORMANCE EVALUATION

LOUISIANA PACIFIC CORP Location CHILCO, ID RTO STACK
Date 10-08 thru 10-15-94 Monitor I.D. USI 500C OPAITY

Confidence Interval = CI ZERO DRIFT

$$t_{.975} = 2.447$$

$$n = 7$$

X_i = difference

$$\text{Confidence Interval} = \frac{t_{.975}}{n\sqrt{n-1}} * \sqrt{n * (\sum_c X_i^2) - (\sum_d X_i)^2}$$

$$= \frac{2.447}{7\sqrt{7-1}} * \sqrt{7 * (0.0884) - (0.0576)} = 0.0663$$

$$\text{Accuracy} = \frac{\frac{b}{\text{Mean of the Differences}} + \text{CI}}{\frac{a}{\text{Mean of Reference Method Values}}} * 100\%$$

(% basis)

$$= \frac{|\underline{\hspace{2cm}}| + \underline{\hspace{2cm}}}{\underline{\hspace{2cm}}} * 100\% = \underline{\hspace{2cm}}\%$$

✓

$$\text{Accuracy} = \frac{b}{\text{Mean of the Differences}} + \text{CI}$$

(conc. basis)

$$= \frac{0.034}{\underline{\hspace{2cm}}} + 0.066 = 0.10$$

32/4

CONTINUOUS EMISSION MONITORING SYSTEM (CEMS)
RELATIVE ACCURACY TEST AUDIT (RATA)

LOUISIANA - PACIFIC CORP Location CHILCO, 10 RTO STACK
 Monitor Manufacturer USI Model No. 500C
 Type of Monitor (SO₂, NO_x, HC, etc.) OPACITY
 Unit of Measurement (% , ppm, wet or dry basis) %
 Emission Standard (provide units) 20%
 Measurement Principle (NDIR, chemiluminescent, etc.) _____
 Monitor System Type (extractive, in-situ) _____
 Range of Instrument _____ Span Gas Concentration _____

RELATIVE ACCURACY TEST <u>SPAN DRIFT</u>						
Test No.	Date	Time	Reference Method ADJ Value	CAL STD CEMS Value	Difference X _i	Difference X _i ²
1	10-08	0834-0900	—	45.94	—	—
2	10-09		45.71		-0.24	0.0576
3	10-10		45.87		-0.07	0.0049
4	10-11		45.79		-0.16	0.0256
5	10-12		45.76		-0.18	0.0324
6	10-13		45.62		-0.32	0.1024
7	10-14		45.83		-0.11	0.0121
8	10-15	↓	45.92	↓	-0.02	0.004
9						
10						
11						
12						
Mean			a		b -0.157	
Sum of the Differences						c 0.2354
(Sum of the Differences) ²					d 1.21	

EXAMPLE CALCULATION OF
CONFIDENCE INTERVAL AND RELATIVE ACCURACY
CONTINUOUS EMISSION MONITORING SYSTEM PERFORMANCE EVALUATION

LOUISIANA-PACIFIC CORP. Location CHILCO, ID RTO STACK
Date 10-08 thru 10-15-94 Monitor I.D. US1 500 C OPACITY

Confidence Interval = CI

SPAN DRIFT

$$t_{.975} = 2.447$$

$$n = 7$$

X_i = difference

$$\text{Confidence Interval} = \frac{t_{.975}}{\sqrt{n-1}} * \sqrt{n * (\sum_c X_i^2) - (\sum_d X_i)^2}$$

$$= \frac{2.447}{\sqrt{7-1}} * \sqrt{7 * (0.2354) - (1.21)} = 0.373$$

$$\text{Accuracy} = \frac{\frac{b}{\text{Mean of the Differences}} + \text{CI}}{\frac{a}{\text{Mean of Reference Method Values}}} * 100\%$$

(% basis)

$$= \frac{|\underline{\hspace{2cm}}| + \underline{\hspace{2cm}}}{\underline{\hspace{2cm}}} * 100\% = \underline{\hspace{2cm}}\%$$

$$\text{Accuracy} = \frac{b}{\text{Mean of the Differences}} + \text{CI}$$

(conc. basis)

$$= |\underline{0.157}| + \underline{0.373} = \underline{0.530}$$

ATO Stack Normal Opacity Calibration Report
LOUISIANA PACIFIC
09:00:00 10-08-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-2.15	0.00	0.00	104.35
Span	45.94	44.49	45.66	-0.28	30.81

Dirt Drift % -3.70

Stack Taper Ratio (Exit/Monitoring) 1.00

=====

Manually CALIBRATED BEFORE THIS CALIBRATION

RTD Stack Normal Opacity Calibration Report
LOUISIANA PACIFIC
09:00:00 10-09-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-3.03	-0.00	-0.00	104.10
Span	45.94	44.50	45.71	-0.24	30.69

Dirt Drift % -3.59

Stack Taper Ratio (Exit/Monitoring) 1.00

=====

RTD Stack Normal Opacity Calibration Report
LOUISIANA PACIFIC
09:00:00 10-10-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-1.78	-0.00	-0.00	103.59
Span	45.94	44.91	45.87	-0.07	30.35

Dirt Drift % -3.35

Stack Taper Ratio (Exit/Monitoring) 1.00

=====

RTO Stack Normal Opacity Calibration Report
LOUISIANA PACIFIC
09:00:00 10-11-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-1.81	0.00	0.00	103.66
Span	45.94	44.80	48.79	-0.16	30.47

Dirt Drift % -3.38

Stack Taper Ratio (Exit/Monitoring) 1.00

=====

RTD Stack Normal Opacity Calibration Report
LOUISIANA PACIFIC .
09:00:00 10-12-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-1.90	-0.00	-0.00	103.83
Span	45.94	44.74	45.76	-0.18	30.54

Dirt Drift % -3.46

Stack Taper Ratio (Exit/Monitoring) 1.00

=====

RTG Stack Normal Opacity Calibration Report

LOUISIANA PACIFIC

08:00:00 10-13-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-1.83	0.00	0.00	103.69
Span	45.94	44.62	45.62	-0.32	30.66

Dirt Drift % -3.39

Stack Taper Ratio (Exit/Monitoring): 1.00

RTO Stack Normal Opacity Calibration Report
LOUISIANA PACIFIC
09:00:00 10-14-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-1.31	0.00	0.00	103.66
Span	45.94	44.35	45.33	-0.11	30.42

Dirt Drift % -3.38

Stack Taper Ratio (Exit/Monitoring) 1.00

=====

RTD Stack Normal Opacity Calibration Report

LOUISIANA PACIFIC

09:00:01 10-15-1994

SAMPLE	CALIBRATION STANDARD	READING	ADJUSTED READING	PERCENT DRIFT	PERCENT TRANSMISSION
Zero	0.00	-1.90	-0.00	-0.00	103.83
Span	45.94	44.89	45.92	-0.02	30.37

Dirt Drift % -3.46

Stack Taper Ratio (Exit/Monitoring) 1.00