

Corrective Action Form

Project Name and Number: Phase 3 Yakima Valley Nitrate Study

Sample Dates Involved: April 2010

Measurement Parameter: All analyses.

Acceptable Data Range: _____

Problem Areas Requiring Corrective Action:

Section 3.1 of the QAPP states: "The laboratories performing the sample analysis of drinking water analytes ...are SDWA certified and/or accredited."

Measures Required to Correct the Problem:

EPA's Manchester Laboratory and Cascade Analytical Laboratory are Safe Drinking Water Act (SDWA) certified for nitrate analysis in drinking water. EPA's Manchester Laboratory used other standard methods for the analysis of metals and pesticides. The UNL Laboratory, USGS Laboratories and Ada Laboratory are not SDWA certified.

Means of Detecting Problems and Verifying Correction:

Initiators Name: Curt Black  Date: 8-1-2012

Project Manager: Curt Black  Date: 8-1-2012

QA Manager:  Gina Groppo-Grove Date: 8-1-12